



Alt-R™ CRISPR-Cas12a (Cpf1) System

Delivery of ribonucleoprotein complexes into Jurkat T cells using the Neon™ Transfection System

> SEE WHAT MORE WE CAN DO FOR YOU AT WWW.IDTDNA.COM.

custom oligos • qPCR • next generation sequencing • RNAi • genes & gene fragments • CRISPR genome editing

REVISION HISTORY

Version	Date released	Description of changes
2	June 2021	Updated to include new Cas12a variants (A.s. Cas12a (Cpf1), A.s. Cas 12a (Cpf1) <i>Ultra</i> , L.b. Cas12a <i>Ultra</i>)
1.1	May 2018	Added note about use of updated Cas12a enzymes (V3)
1	May 2017	Original protocol

Table of contents

Revision history	2
Introduction	4
Important	4
Consumables and equipment	5
Prepare cell cultures for electroporation	5
Prepare the crRNA	6
Form the RNP complex	6
Prepare Neon Transfection System	6
Perform electroporation of cells	6
References	8

INTRODUCTION

This protocol describes the optimized delivery of a Cas12a (Cpf1) ribonucleoprotein (RNP) complex, containing Alt-R™ CRISPR-Cas12a (Cpf1) crRNA and any of the available Cas12a variants (A.s. Cas12a (Cpf1), A.s. Cas 12a (Cpf1) *Ultra*, L.b. Cas12a *Ultra*), into Jurkat T cells using the Neon™ Transfection System (Thermo Fisher) with 10 µL electroporation cuvettes [1]. Application of this protocol can be found in a [DECODED newsletter article](#) [2]. Other cell lines may require optimization of electroporation parameters and RNP concentration to ensure desired editing efficiency for your experimental needs. To investigate on-target mutations and estimate editing efficiency, we recommend using the Alt-R Genome Editing Detection Kit [3].

The CRISPR-Cas12a (Cpf1) system is distinct from the more commonly used CRISPR-Cas9 system. For example, Cas12a (Cpf1) nuclease does not require a tracrRNA. This nuclease recognizes a T-rich, protospacer-adjacent motif (PAM: TTTV, where V is an A, C, or G base) and creates a staggered double-stranded DNA cut with a 5' overhang. For additional information, visit www.idtdna.com/CRISPR-Cpf1.

IMPORTANT

1. **Use low-passage, healthy cells.** A critical factor affecting the success of electroporation is the health of the cells. It is important to:
 - Use the lowest passage number cells available
 - Subculture cells for at least 2–3 days before the electroporation procedure
 - Replace the media the day before electroporation
 - Determine the optimal confluency for your cell type
2. Wash cells before electroporation. FBS and trypsin may contain RNase activity that can quickly degrade CRISPR RNA components. Therefore, it is crucial to wash cells with PBS to remove FBS-containing media or trypsin. Alternatively, for adherent cells, use “enzyme-free” dissociation media instead of trypsin.
3. Alt-R Cas12a (Cpf1) Electroporation Enhancer, a non-targeting, single-stranded carrier DNA, is required for successful electroporation. See the Performance section at www.idtdna.com/CRISPR-Cpf1 for supporting data.
4. We recommend the use of appropriate controls in your experiment such as HPRT-specific, positive control crRNA and a non-targeting negative control. For suggested sequences for studies in human, mouse, or rat cells, see [page 5](#).

To detect on-target editing of the suggested HPRT control crRNA, Alt-R HPRT PCR Primer Mix (available for human, mouse, or rat) may be used with the Alt-R Genome Editing Detection Kit (T7 endonuclease I assay).

CONSUMABLES AND EQUIPMENT

Instruments, kits, and reagents	Ordering information
Neon Transfection System	Thermo Fisher Scientific (Cat # MPK5000)
Neon Transfection System 10 µL Kit	Thermo Fisher Scientific (Cat # MPK1096)
RPMI-1640 Medium (RPMI)	ATCC (Cat # 30-2001)
Fetal bovine serum (FBS)	General laboratory supplier
1X Phosphate buffered saline (PBS)	General laboratory supplier
Alt-R CRISPR-Cpf1 crRNA	IDT custom crRNA (www.idtdna.com/CRISPR-Cpf1)
Alt-R Cas12a (Cpf1)	
<ul style="list-style-type: none"> variant A.s. Cas12a (Cpf1) 	1081068, 1081069
<ul style="list-style-type: none"> variant A.s. Cas 12a (Cpf1) <i>Ultra</i> 	10001272, 10001273, 10007804
<ul style="list-style-type: none"> variant L.b. Cas12a <i>Ultra</i> 	10007922, 10007923, 10007924
Nuclease-Free IDTE, pH 7.5 (1X TE solution)	IDT (Cat # 11-01-02-02)
	IDT (Cat # 1076300)
Alt-R Cas12a (Cpf1) Electroporation Enhancer	 Note: The electroporation enhancer does not have significant homology to the human, mouse, or rat genomes, and has been tested as carrier DNA in multiple cell lines, including HEK-293, Jurkat, and HeLa.

Improved enzymes

The Alt-R Cas12a (Cpf1) enzyme has recently been further optimized. The latest version (V3) can be directly substituted into this protocol in place of the prior Alt-R Cpf1 enzyme.

To order control crRNAs, enter the appropriate sequence into the Cas12a crRNA ordering tool (accessible at www.idtdna.com/CRISPR-Cpf1). These sequences are available online for copying and pasting into the ordering tool.

Recommended control crRNAs	Sequence
Positive control Cas12a crRNA, Human <i>HPRT1</i>	GGTTAAAGATGGTTAAATGAT
Positive control Cas12a crRNA, Mouse <i>Hprt</i>	GGATGTTAAGAGTCCCTATCT
Positive control Cas12a crRNA, Rat <i>Hprt1</i>	ATGCTTAAGAGGTATTTGTTA
Negative control Cas12a crRNA #1	CGTTAATCGCGTATAATACGG
Negative control Cas12a crRNA #2	CATATTGCGCGTATAGTCGCG
Negative control Cas12a crRNA #3	GGCGCGTATAGTCGCGCGTAT

Prepare cell cultures for electroporation

- Do not use freshly thawed cells for electroporation: passage your cells 1X after thawing, and verify that they grow well and look healthy.
- Use cells with the lowest passage number possible.
- Change the cell culture media on the cells 1 day before electroporation.
- Split cells, if necessary, to obtain optimal confluency for electroporation.



Note: For Jurkat cells, optimal cell density is between 1×10^5 and 1×10^6 cells/mL at the time of transfection [4].

Prepare the crRNA

1. At first use, resuspend Alt-R CRISPR-Cas12a crRNA in IDTE Buffer to create a 100 μM stock solution. For assistance, use the IDT Resuspension Calculator at www.idtdna.com/SciTools.

Store resuspended RNAs at -20°C .

2. Dilute crRNA (from **Step 1**) to a 75 μM working dilution. You will need 1 μL of working dilution for each electroporation in **Form the RNP complex, Step 1**.

Form the RNP complex

1. For each well undergoing electroporation, combine the crRNA and Cas12a (Cpf1) Nuclease components, gently swirling the pipet tip while pipetting:

Component	Amount
Alt-R CRISPR-Cas12a crRNA (from Prepare the crRNA, Step 2)	1.0 μL
Alt-R Cas12a (Cpf1) Nuclease	1.0 μL
Total volume	2.0 μL*

* This 2 μL volume is for 1 electroporation reaction; scale up as necessary for your experiment—we recommend making 1.2X volume needed to correct for pipetting errors.

2. Incubate the mixture at room temperature for 10–20 min.

Prepare Neon Transfection System

1. Turn on the Neon system.
2. Enter electroporation settings, or choose a setting from the optimization protocol.



Note: In internal experiments, the optimum settings for Jurkat cells were found to be 1600 V, 10 ms pulse width, 3 pulses [2].

3. Set up the Neon Pipette Station by filling the Neon Tube with Electrolytic Buffer (included in the Neon Transfection System Kit) and insert into the station.

Perform electroporation of cells

1. Prepare the Alt-R Cas12a (Cpf1) Electroporation Enhancer. For assistance, use the Resuspension Calculator at www.idtdna.com/SciTools.
 - a. At first use, resuspend the Alt-R Cas12a (Cpf1) Electroporation Enhancer to 100 μM in IDTE to create a stock solution.
 - b. For each set of experiments, dilute stock to 10.8 μM (working solution). You will need 2 μL of working solution for each electroporation in **Step 11**.
2. Prepare culture plates to receive cells following electroporation.
 - a. Set 1: For each electroporation sample, fill a well with 190 μL of culture media (RPMI, 10% FBS) to resuspend cells in **Step 16**.
 - b. Set 2: For each electroporation sample, fill 3 wells with 150 μL of culture media (RPMI, 10% FBS) for growth in **Step E17**.
 - c. Store all plates in a tissue culture incubator (37°C , 5% CO_2).
3. Pipette cells up and down to dissociate cell clumps.

4. Remove as much supernatant as possible without disturbing the pellet.
5. Wash cells in 5 mL of 1X PBS, and then centrifuge at 600 rpm (Beckman GH-3.8 rotor: 58 x g) for 10 min at room temperature.



Note: FBS in the growth media may contain RNase activity. Therefore, it is crucial to wash the cells with PBS after spinning down.

6. Remove as much supernatant as possible without disturbing the pellet.
7. Resuspend cells by adding 8 μL of Resuspension Buffer R per 5×10^5 cells.
8. For each electroporation, add the following to a 200 μL PCR tube:

Component	Amount (μL)*
crRNA:Cas12a RNP complex (from Form the RNP complex, Step 2)	2
Cell suspension (from Step 10)	8
10.8 μM Alt-R Cas12a (Cpf1) Electroporation Enhancer (from Step 1b)	2
Total volume	12

* The final concentration for each electroporation is 6.25 μM crRNA, 5.25 μM Cas12a nuclease, and 1.8 μM Cas12a (Cpf1) electroporation enhancer.

9. Insert a Neon Tip into the Neon Pipette.
10. Pipette 10 μL of cell/RNP complex mixture (from [Step 11](#)) into the Neon Tip, avoiding air bubbles.
11. Insert the Neon Pipette and Tip into the Pipette Station. Verify the presence of Electrolytic Buffer in the Neon Tube.
12. Press **Start**.
13. After electroporation, transfer cells to wells containing 190 μL of pre-warmed culture media (RPMI, 10% FBS) (from [Step 2a](#)) and slowly resuspend.
14. Transfer 50 μL of resuspended cells in triplicate to the wells containing 150 μL of pre-warmed culture media (RPMI, 10% FBS) (from [Step 2b](#)).
15. Incubate cells in a tissue culture incubator (37°C, 5% CO₂) for 72 hr.

To investigate on-target mutations with the Alt-R Genome Editing Detection Kit (T7EI mismatch assay), follow the protocol [\[3\]](#).

REFERENCES

1. (2014) Neon Transfection System for transfecting mammalian cells, including primary and stem cells, with high transfection efficiency. Thermo Fisher. Available at https://tools.thermofisher.com/content/sfs/manuals/neon_device_man.pdf. (Accessed June 28, 2021)
2. Turk R and Prediger E. (2016) Successful CRISPR genome editing in hard-to-transfect cells. Available at <https://www.idtdna.com/pages/education/decoded/article/successful-crispr-genome-editing-in-hard-to-transfect-cells>. (Accessed June 28, 2021)
3. (2017) Alt-R Genome Editing Detection Kit protocol. Integrated DNA Technologies. Available at <https://sfvideo.blob.core.windows.net/sitefinity/docs/default-source/protocol/alt-r-genome-editing-detection-kit.pdf>. (Accessed June 28, 2021)
4. American Type Culture Collection (ATCC) .<https://www.atcc.org/products/tib-152>

Alt-R™ CRISPR-Cas12a (Cpf1) System

Delivery of ribonucleoprotein complexes into Jurkat T cells using the Neon™ Transfection System

Technical support: applicationsupport@idtdna.com

For more than 30 years, IDT's innovative tools and solutions for genomics applications have been driving advances that inspire scientists to dream big and achieve their next breakthroughs. IDT develops, manufactures, and markets nucleic acid products that support the life sciences industry in the areas of academic and commercial research, agriculture, medical diagnostics, and pharmaceutical development. We have a global reach with personalized customer service.

> SEE WHAT MORE WE CAN DO FOR YOU AT WWW.IDTDNA.COM.

For Research Use Only. Not for use in diagnostic procedures. Unless otherwise agreed to in writing, IDT does not intend these products to be used in clinical applications and does not warrant their fitness or suitability for any clinical diagnostic use. Purchaser is solely responsible for all decisions regarding the use of these products and any associated regulatory or legal obligations.

© 2021 Integrated DNA Technologies, Inc. All rights reserved. Trademarks contained herein are the property of Integrated DNA Technologies, Inc. or their respective owners. For specific trademark and licensing information, see www.idtdna.com/trademarks.
MAPPS Doc ID# RUO21-0214_001 08/21